


AL 285 **Logger-Check** For thermal validation *DAC UNIVERSAL*

In combination with the software Winlog.validation the AL 285 offers you a simple, on site functional check of the measuring channels.



EN ISO 17665-1, paragraph 9.1.4

Verification of the calibration value of the measuring instrument that is used for the validation of the sterilization process at process condition.

- Easy and reproducible test with automatic generation of reports.
- The tests are stored permanently in the database

